

High-numerical-aperture cryogenic light microscopy for increased precision of superresolution reconstructions

Marc Nahmani^{a,1}, Conor Lanahan^{a,2}, David DeRosier^a, and Gina G. Turrigiano^{a,3}

^aDepartment of Biology, Brandeis University, Waltham, MA 02454

Contributed by Gina G. Turrigiano, March 1, 2017 (sent for review November 3, 2016; reviewed by Sam Hess and Cindi Schwartz)

Superresolution microscopy has fundamentally altered our ability to resolve subcellular proteins, but improving on these techniques to study dense structures composed of single-molecule-sized elements has been a challenge. One possible approach to enhance superresolution precision is to use cryogenic fluorescent imaging, reported to reduce fluorescent protein bleaching rates, thereby increasing the precision of superresolution imaging. Here, we describe an approach to cryogenic photoactivated localization microscopy (cPALM) that permits the use of a room-temperature high-numerical-aperture objective lens to image frozen samples in their native state. We find that cPALM increases photon yields and show that this approach can be used to enhance the effective resolution of two photoactivatable/switchable fluorophore-labeled structures in the same frozen sample. This higher resolution, two-color extension of the cPALM technique will expand the accessibility of this approach to a range of laboratories interested in more precise reconstructions of complex subcellular targets.

superresolution | single molecule | localization microscopy | photoactivated localization microscopy | cryogenic microscopy

Single-molecule localization-based microscopy (SMLM) techniques have enhanced our ability to resolve the intricate spatial relationships between identified subcellular proteins (1–3). These techniques routinely achieve an average localization precision of ~25 nm (4). Recent SMLM studies have begun to fundamentally alter our understanding of cytoarchitecture (5) and the dynamic interplay of cellular proteins (6, 7).

Despite this progress, many subcellular structures remain difficult to study because they are composed of a compact network of single-molecule-sized elements. These structures lie beyond current SMLM approaches because the accuracy of SMLM techniques is dependent on the achieved precision in localizing individual fluorophores, which primarily relies on the number of photons collected from each of these molecules (8). Unfortunately, genetically encoded probes visualized at room temperature suffer from relatively low quantal yield and high numbers of unactivatable fluorophores, which limit localization precision and the completeness of SMLM reconstructions (9), respectively. Equally lacking are localization methods based on antibody labeling because antibodies necessarily add ~10 nm of uncertainty in localizing the molecules of interest (10).

To study in detail the native structural relationships of small, densely packed proteins, it would be ideal if one could (*i*) study frozen, hydrated specimens preserved in their native state, (*ii*) increase the localization precision of individual molecules to resolve the intricate relationships of small, dense structures and more faithfully reconstruct these objects, and (*iii*) combine these localization maps of identified objects with the ultrastructural information gained via cryogenic electron microscopy (cryoEM). Examples of the combined use of a cryogenic SMLM method, cryogenic photoactivated localization microscopy (cPALM), and cryoEM have illustrated the utility of combining these methods. However, these studies used cryogenic stages with air-coupled

objective lenses that consequently limited the collection of photons and the effective resolution (11–13). Here we developed a cryogenic stage and objective interface that extends the cPALM approach by using a commercially available high-numerical-aperture (N.A.) lens (≥ 1.2 N.A.), enabling more precise fluorescent imaging of frozen (–140 to –150 °C) hydrated specimens and offering significant improvement over room-temperature imaging with identical equipment. Unlike the only other high-N.A. implementation of cPALM in which the objective is held at temperatures below –135 °C (14), we developed an approach that keeps the objective at ambient temperature and therefore can be used on a conventional microscope. Moreover, we find that like photoactivatable GFP [PaGFP (11)], the photoswitchable protein PSmOrange (15) is able to switch at cryogenic temperatures, adding to the repertoire of photoswitchable proteins capable of switching at cryogenic temperatures (13), and enabling simultaneous localization and comparison of two photoactivatable/switchable proteins within the same sample under cryogenic conditions. These samples can then be imaged using cryoEM to correlate SMLM localization “maps” of identified molecules with ultrastructural information (11, 12).

To image frozen hydrated samples with a short working distance high-N.A. lens at temperatures below the threshold for ice crystal formation (≤ -135 °C), we built a cryogenic stage that is cooled by the flow of cold nitrogen gas and that maintains samples on electron microscopy grids at stable cryogenic temperatures (Fig. 1 *D* and *E*). Samples were optically coupled to the coverslip by a cryofluid that freezes below –150 °C and does not boil at ambient temperature

Significance

The precision of localization-based superresolution microscopy techniques fundamentally relies on the point-spread function of the optical system and the number of photons one can collect. Here, we report that by using a high-numerical-aperture objective lens and a custom cryogenic stage, we are able to increase photon yield by 2–3-fold over room temperature, thereby achieving more precise superresolution reconstructions of complex subcellular structures.

Author contributions: M.N., D.D., and G.G.T. designed research; M.N., C.L., and D.D. performed research; M.N., C.L., D.D., and G.G.T. analyzed data; and M.N., D.D., and G.G.T. wrote the paper.

Reviewers: S.H., University of Maine; and C.S., National Institute of Allergy and Infectious Diseases/NIH.

Conflict of interest statement: The authors built successive versions of the cPALM stage and objective collar and eventually patented the current system (Publ US 2015/0248002 A1). Design elements and detailed scaled images for constructing a similar cryogenic stage and objective collar system can be found in this patent, as well as in the annotated computer-aided design (CAD) drawings available as [Movies S1](#) and [S2](#).

¹Present address: Division of Sciences and Mathematics, University of Washington Tacoma, Tacoma, WA 98402.

²Present address: Massachusetts General Hospital, Boston, MA 02114.

³To whom correspondence should be addressed. Email: turrigiano@brandeis.edu.

This article contains supporting information online at www.pnas.org/lookup/suppl/doi:10.1073/pnas.1618206114/-DCSupplemental.




Fig. 1. Cryogenic cold stage for single-molecule localization microscopy. (A) CAD drawing and (B and C) images of high-N.A. objective lens fitted with a collar to run immersion fluid across the lens to prevent lens freezing. (D) CAD drawing of cryogenic cold stage, displaying inlet and outlet ports that run cooled nitrogen gas to a 3-mm sample holder and upward toward objective lens (Movies S1 and S2). (E) Higher magnification view of stage sample holder showing orientation of sample to cryofluid and coverslip. (F) Simulation of thermodynamics when running immersion fluid past the objective lens that is optically coupled via a cryogenic liquid to a sample cooled to -145°C . (G) Raw temperature data read from a thermocouple embedded in the copper rod directly under the sample, and another placed at RT. Data are from a representative cold run on a sample of HEK 293 cells. After cooled nitrogen gas is turned on (arrow), the copper sample holder cools down to steady state (-143 to -145°C) within 20 min and can maintain that temperature as long as a supply of liquid nitrogen is available. (Inset) Temperature readout from 45–75 min after the stage began cooling, showing that temperature is stable at steady state to within $\pm 1^{\circ}\text{C}$.

(Fig. 1E). To prevent the immersion fluid between the coverslip and objective from freezing, we developed an objective lens collar that runs room-temperature (RT) immersion fluid past the end of a commercially available water-immersion objective lens (Fig. 1A–C). In this configuration, a temperature drop of $\sim 150^{\circ}\text{C}$ occurs over a distance of a few hundred micrometers, allowing cryogenic samples to be kept below temperatures for ice crystal formation while the objective lens is maintained at a safe operating temperature (~ 15 – 18°C) (Fig. 1F and Table S1). The cold stage achieves stable cryogenic temperatures within 15–20 min and can maintain a desired sample temperature during imaging within $\pm 1^{\circ}\text{C}$ indefinitely, given a steady supply of cooled gas (Fig. 1G).

One of the reported benefits of cryogenic fluorescent imaging is the relative reduction in fluorophore bleaching, theoretically increasing photon yield and thus localization precision per

fluorescent molecule (14, 16). To quantify this potential increase in photon yield, we first characterized the activation and bleach kinetics of known [PaGFP (11)] and untested [PSmOrange (15)] cryogenic photoactivatable/photoswitchable purified proteins at RT (21 – 26°C) and cryogenic temperature (-140°C) (Fig. 2A and B). Whereas the rate of PaGFP activation at -140°C was comparable to that at RT [$P > 0.05$, Kolmogorov–Smirnov (KS) test], PSmOrange showed a trend toward activating more slowly, taking longer to reach asymptote at -140°C versus RT ($P < 0.07$, KS test; Fig. 2C). However, both PaGFP and PSmOrange had significantly reduced rates of bleaching and significantly more captured photons at -140°C versus RT (~ 2.1 -fold increase for PaGFP and 1.5–2-fold for PSmOrange; $P < 0.001$ for both conditions, KS test, $n \geq 3$ per condition; Fig. 2C and Fig. S1).

To determine whether bleach protection at cryogenic temperature results in increased photon yield and localization precision of individual molecules, we performed SMLM on PaGFP protein diluted on electron microscopy (EM) grids at RT, and then performed cPALM on the identical samples. We found that individual photoactivated PaGFP molecules imaged at -140°C were ~ 2.6 -fold brighter (median photon count = 414.5 ± 68.7 vs. 157.6 ± 11.8 , -140°C vs. RT, respectively, $P = 0.0001$, Mann–Whitney U test, $n = 3$ samples per condition), lasted 2.1-fold longer (median frames detected = 3.4 ± 0.2 vs. 1.6 ± 0.07 frames, $P = 0.00001$, Mann–Whitney U test), and were localized with greater precision (15% decrease in error, median uncertainty of localization = 20.9 ± 0.02 vs. 24.1 ± 0.9 nm, $P = 0.00001$, Mann–Whitney U test; Fig. 2D and E). Interestingly, in analyzing identical-sized areas with the same imaging parameters, there were ~ 5.3 -fold more localized molecules when imaging at -140°C compared with RT (mean total points detected = $2,854 \pm 233$ vs. 536 ± 28 , $P = 0.01$, unpaired two-tailed t test). Together, these results demonstrate that cPALM imaging results in a marked improvement in the photon yield and localization precision of individual fluorophores, and provides the unexpected benefit of allowing for more faithful reconstructions by increasing the number of detectable photoactivated molecules.

To examine whether these advantages generalized to cPALM imaging of complex subcellular structures, we grew HEK 293 cells on EM grids and then doubly transfected them with Utrophin–PaGFP and Vimentin–PSmOrange. Utrophin is an F-actin binding protein and can act as a reporter for polymerized actin filaments (17), whereas vimentin is an intermediate filament protein expressed in a complementary filamentous network (15). We next imaged these transfected cells using either PALM at RT or cPALM at -140°C on the identical EM grids. Consistent with our single-molecule analyses of purified protein, we found that imaging these cytoskeletal networks using cPALM resulted in increased photon yields, an increase in the number of localized molecules per exposure, and greater precision in localizing and reconstructing these subcellular structures (Fig. 3A and Fig. S2). Specifically, imaging at cryogenic temperature resulted in a 2.4–2.7-fold increase in photon yield per detected molecule and a commensurate 31–32% reduction in single-molecule localization error (33.8 ± 2.0 vs. 44.2 ± 2.4 nm, and 33.9 ± 1.9 vs. 44.6 ± 2.5 nm, $P = 0.03$ for PaGFP and PSmOrange, cPALM vs. RT respectively, unpaired two-tailed t test) versus when imaging adjacent cells at RT (Fig. 3C).

We reasoned that one method to verify that the reduction in single-molecule localization error resulted in a more precise cytoskeletal reconstruction would be to analyze the PaGFP and PSmOrange pixel overlap in regions across the cells where the two signals were in close proximity (i.e., $\leq 1 \mu\text{m}$). We found that the correlation coefficient for overlapped pixels ($-1 =$ no signal overlap, $+1 =$ complete signal overlap) in these regions was significantly reduced (nearly 10-fold) when imaging at cryogenic temperature versus RT, indicating greater precision in resolving these two signals as evinced by the enhanced separation between the two complementary cytoskeletal reconstructions ($R = 0.02 \pm 0.04$ vs. 0.2 ± 0.03 ,




Fig. 2. Bleach protection, increased single-molecule photon yield, and reduced localization uncertainty of purified protein imaged at cryogenic temperatures. (A) Images of purified PaGFP protein (10 μ M; wide-field microscope with commercial cryogenic stage) after activation at RT (Top) and at -140 $^{\circ}$ C (Bottom). Note the slight increase in background signal at cryogenic temperatures. (B) Images of purified PSmOrange protein (40 μ M, confocal microscope with commercial cryogenic stage) after activation at RT (Top) and cryogenic temperatures (Bottom). Red is the activated/switched state (Left), yellow/orange is the unactivated/unswitched state (Right). Note the clear activation in the red channel, and absence of fluorescence after switching in the yellow channel, in both RT and cryogenic conditions. (C) Activation (Top) and bleach (Bottom) rates for purified PaGFP (10 μ M, Left) and PSmOrange (40 μ M, Right). Activation graphs show the rate of activation from baseline to maximal activation, normalized for each individual run. Shaded regions around curves indicate SEM. Bleach curves represent the rate that activated/switched areas were bleached. (D) Scatter plot of purified PaGFP single-molecule data showing increased photon yield and detected single-molecule persistence at cryogenic temperatures (triangles) vs. RT (circles). (E) Radar plot quantifying single-molecule PaGFP data as in D, showing photon yield (mean and median), localization uncertainty, and the number of localized PaGFP molecules over a given area and exposures when imaging at RT vs. cryogenic temperatures. [Scale bars: 100 μ m for A (all panels) and B (all panels).]

$P = 0.02$, unpaired two-tailed t test, $n = 3$ cells per condition; Fig. 3B and Fig. S3).

To verify that our cPALM approach is compatible with cryoEM, we generated cryoEM images of tobacco mosaic virus that had been treated with our cryofluid (3-methyl-1-pentene) at -145 $^{\circ}$ C and then rinsed with cold liquid ethane at -145 $^{\circ}$ C to elute the cryofluid before cryoEM imaging. When comparing these

samples with identical samples that had been rinsed in ethane but not exposed to the cryofluid, we found that both treated and untreated samples showed reflections to 1.2-nm resolution (Fig. S4). Hence, contact with the cryofluid used for cPALM imaging does not alter the specimen at ~ 1 -nm resolution.

Interestingly, single-molecule analysis of purified PaGFP protein revealed a 2.6-fold increase in photon yield but only a




Fig. 3. Improved precision with cryogenic single-molecule localization microscopy. (A) HEK 293 cells transfected with Utrophin-PaGFP (actin, green) and Vimentin-PSmOrange (intermediate filaments, red). Cells imaged at RT or cryogenic temperatures were always compared with adjacent cells grown on the identical EM grid. (B) Magnified segments from white boxes shown in merged images, showing localization precision (as a function of plotted molecule size) and amount of colocalization for RT (Top) and cryogenic (Bottom) conditions. Pearson's r coefficients on bottom of images report the average pixel overlap ($-1 =$ no overlap, $+1 =$ complete overlap) for colocalized red and green pixels within regions of interest across the entire cell ($n = 3$ cells per condition). (C) Box plots displaying photon yield per molecule and localization uncertainty for Utrophin-PaGFP and Vimentin-PSmOrange at RT and -140 $^{\circ}$ C. Black bars indicate medians of distributions. (D) Box plots of Pearson's r coefficients for colocalized red/green pixels in reconstructions from HEK cells at RT and -140 $^{\circ}$ C. (Scale bars: A, 5 μ m, applies to all panels within each condition; B, 200 nm for top and bottom panels.)

15% decrease in localization uncertainty. This potential discrepancy between theoretical and observed uncertainty is due to a relatively small increase in the Gaussian half-width (~1.2-fold) and a doubling of the SD of background signal observed when imaging purified protein at cryogenic temperatures, possibly due to temperature-induced changes in focus that can be attributed to the slow cooling of the foam insulation of the stage. Combined, these factors fully accounted for the observed localization uncertainty of purified PaGFP molecules (Table S2). In contrast to the purified protein analyses, the observed Gaussian half-widths (1.2-fold larger) and background signals (1.2–1.6-fold larger) of Utraphin–PaGFP and Vimentin–PsmOrange at cryogenic temperature were only slightly higher than at RT. Accordingly, our observed 2.4–2.7-fold increase in photon yield in these cytoskeleton reconstruction experiments translated to a 31–32% decrease in localization uncertainty at –140 °C (Table S2).

Previous reports of cryogenic SMLM techniques have used low-N.A. objective lenses to demonstrate that these tools can effectively couple information about molecular identity and localization with structural information from electron micrographs (11–13). In addition to these benefits, here we find that by using a high-N.A. lens the cPALM method produces more faithful and precise SMLM reconstructions that will enable higher resolution molecular maps of dense subcellular proteins. Importantly, the cPALM stage and objective collar can be customized to fit commercial upright fluorescent microscopes and commercially available high-N.A. objective lenses (see *Methods* and *Movies S1* and *S2*). We hope this higher resolution extension of the cPALM technique will expand the accessibility of this approach to a range of laboratories interested in precise reconstructions of complex subcellular targets.

Methods

Photoactivatable Constructs, Protein Purification, and Transfections. Photoactivatable proteins and tagged constructs were purchased from Addgene: pRSETA-PaGFP was a gift from Jennifer Lippincott-Schwartz, Janelia Research Campus, Ashburn, VA (Addgene plasmid #11911), UtrCH-PaGFP was a gift from William Bement, University of Wisconsin-Madison, Madison, WI (Addgene plasmid #26738), and pBAD-PsmOrange and pVimentin-PsmOrange were gifts from Vladislav Verkhusha, Albert Einstein College of Medicine, Bronx, NY (Addgene plasmid #31886). To purify photoactivatable proteins, a single bacterial colony grown from an Addgene stab was picked from a fresh agar plate, and then grown in the presence of 20% arabinose and the appropriate selective antibiotic until logarithmic growth was achieved. Protein production was induced with 1 mM isopropyl β -D-1-thiogalactopyranoside and bacterial cultures were then allowed to incubate overnight in a 37 °C shaker. Next, bacterial cultures were centrifuged, lysed (CellLytic B Plus kit, Sigma, CB0050), and then purified using a His-nickel gravity column kit (Clontech 636658). Concentrated proteins were embedded in an 11–13% polyacrylamide gel for subsequent imaging at RT and cryogenic temperatures. Imaging of concentrated proteins at RT and –140 °C was performed using an air-coupled 10 \times lens and a commercially available cryogenic stage (Instec, Inc). For SMLM imaging of PaGFP, a small dab of 5 μ M PaGFP protein in polyacrylamide gel was spread evenly onto a gold or molybdenum EM grid, and a 1.5- μ L drop of 1:1,500 diluted Tetraspeck beads was placed on the sample and allowed to dry. These samples were then imaged using our cPALM stage. HEK 293 cells were doubly transfected with UtrCH-PaGFP and pVimentin-PsmOrange using polyethyleneimine (Polysciences, Inc., #23966) at a 2.5 μ L: 1 or 1.5 μ g DNA ratio. After 16–18 h, transfection media was replaced with fresh Dulbecco's modified Eagle medium and HEK cells were placed in a 37 °C incubator for 2–3 d before imaging. All cells (for RT and –140 °C experiments) were fixed in 4% paraformaldehyde in PBS (for 10 min) before imaging to control for any differences in precision or colocalization related to fixative-based preservation. However, to maximize ultrastructural preservation, most biological applications will likely benefit from using fast-frozen, unfixed samples.

cPALM Stage and Objective Collar Construction. Before stage and objective collar construction, we first determined that we could theoretically maintain our sample below the threshold for ice crystal formation while imaging with a high-N.A. water immersion objective lens maintained at RT. We calculated

these values based on the physics of heat flow and using COMSOL software to model the steady-state thermodynamics with our design materials (Table S1 and Fig. 1F). Once determining that these design constraints were feasible, we built successive versions of the cPALM stage and objective collar and eventually patented the current system (Publ US 2015/0248002 A1). Design elements and detailed scaled images for constructing a similar cryogenic stage and objective collar system can be found in this patent, as well as in the annotated computer-aided design (CAD) drawings available as *Movies S1* and *S2*.

Optical Setup and Operation. The principal elements of our microscope include three linearly polarized lasers: a 473-nm diode laser (103 mW, Laserglow LRS-0473-PFM-00050-05), a 643-nm diode laser (53 mW, CrystaLaser DL640-050), and a 405-nm diode laser (50 mW, Coherent 1069413/AT). Lasers were aligned and directed up to a rigidly supported vertical breadboard that housed all of the optical elements and camera. All lasers were fed into a computer-controlled motorized filter wheel (Thorlabs FW103) altered to operate perpendicular to the vertical breadboard. Lasers were directed 90° downward toward the back aperture of the objective lens by appropriate dichroic lenses, and matching notch filters restricted fluorescence emission to the camera. Fluorescent images were captured by a cooled back-illuminated EMCCD (Andor, DU-897E) with a 512 \times 512 16- μ m² pixel detector. Our measured laser power density ranges at the sample were as follows: 473-nm laser (readout: 80–150 W/cm², activation: 10–18 W/cm²), 643-nm laser (readout: 108–355 W/cm², activation: 0.8–12.7 W/cm²). We used 0.1-s exposures for readout beams and between 0.01- and 0.5-s exposures to photoactivate our samples with either the 405- or 473-nm lasers.

To maintain optical coupling of the objective lens and the specimen at \leq –135 °C, we needed a cryofluid with a refractive index close to that of water, a boiling point well above –23 °C (the calculated temperature at the bottom of the coverslip as shown in Table S1), and a freezing point below –140 °C. Listed below are the three we tested:

Cryofluid	Freezing point, °C	Boiling point, °C	Refractive index
3-methyl-1-pentene	–154	+54	1.38
4-methyl-1-pentene	–153	+54	1.38
70% methanol: 30% 1-propanol	~–148	>+25	~1.38

We primarily used the 70:30% methanol:1-propanol mixture as it had superior surface tension and was less volatile than the pentene fluids. In addition, we used 70% ethanol in dH₂O as an objective immersion fluid (between the objective lens and attached coverslip) because we found that dH₂O alone tended to freeze when using the relatively slow peristaltic pumping speeds required to prevent pressure-induced image artifacts.

We imaged in one of two ways: (i) we placed the EM sample grid on the stage, filled the reservoir with coupling fluid, and then proceeded to focus down on the sample as the stage reached –140 °C, or (ii) in experiments aimed at preserving native ultrastructure we kept the sample under LN₂ until the stage cooled down to –140 °C and then secured the sample on the stage. We then poured some LN₂ into the sample tray and dropped some cryofluid into the tray. Next we carefully picked up the now-frozen drop of cryofluid from the LN₂ in the sample tray, placed it over the sample, and then lowered the objective lens and collar unit just above the cryofluid to cool and dry the coverslip to cryogenic temperature. Once the coverslip, cryofluid, and sample had reached –140 °C, we lowered the objective lens onto the sample and began imaging. Sample temperature was continually monitored with a thermocouple embedded in the copper sample rod, directly under the sample. If the sample temperature fell lower than –148 °C (cryofluid freezing point) or above –135 °C (ice crystal formation threshold), imaging runs were aborted and data were excluded from our analyses.

SMLM Acquisition and Analysis. We wrote custom LabVIEW software routines to control and synchronize all aspects of our image acquisition, including all of the lasers, filter sets, shutters, EMCCD, and A/D devices in our system. We used the ImageJ plugin ThunderStorm (18) to localize, fit, and reconstruct our RT PALM and cPALM data. ThunderStorm is an open-source platform that uses published, well-characterized methods to localize and fit single molecules and calculate uncertainty (Table S2). Briefly, ThunderStorm fits a peak using an image-filtering algorithm (in our case a Wavelet filter with B-spline order = 3 and scale = 2) within a certain threshold [we used 1.2*SD

